

Calibration certificate – additional Annex Werkskalibrierschein – zusätzlicher Anhang



xxxxx
2023xxx

Frequency measurement

Frequenzmessung

Measuring equipment

Messmittel

Description <i>Beschreibung</i>	Equipment No. <i>Prüfmittel-Nr.</i>	Calibration Date <i>Kalibrierdatum</i>
Univ. counter HM8 123 3GHz Rhode & Schwarz	22/138	05.10.2022

Performance check

Funktionskontrolle

- Measured values within specification** *Messwert(e) innerhalb der Spezifikation*
- Measured values beyond specification** *Messwert(e) außerhalb der Spezifikation*

Measurement results

Messergebnisse

reference value [kHz]	reading [kHz]	deviation [kHz]
60	60,000001	-0,000001
120	120,000002	-0,000002
240	240,000005	-0,000005
480	480,000003	-0,000003
960	960,000013	-0,000013