

# Calibration certificate – additional Annex

## Werkskalibrierschein – zusätzlicher Anhang



xxxxx
2023xxx

### Conductivity reference standards – single values

*Referenznormale Leitfähigkeit - Einzelwerte*

### Serial number (code)

*Serien-Nr. (Code)*

L2219050\_230811000

### Measurement results

*Messergebnisse*

measuring number	frequency [kHz]				
	60	120	240	480	960
	Reading [MS/m]	Reading [MS/m]	Reading [MS/m]	Reading [MS/m]	Reading [MS/m]
1	17,101	17,109	17,118	17,132	17,107
2	17,091	17,109	17,119	17,131	17,109
3	17,102	17,102	17,118	17,129	17,104
4	17,11	17,112	17,116	17,132	17,101
5	17,101	17,103	17,121	17,131	17,112
6	17,093	17,115	17,122	17,1233	17,104
7	17,096	17,14	17,119	17,129	17,115
8	17,102	17,103	17,121	17,131	17,141
9	17,103	17,103	17,13	17,1197	17,102
10	17,101	17,104	17,121	17,129	17,109
<b>AV [MS/m]</b>	<b>17,10</b>	<b>17,11</b>	<b>17,12</b>	<b>17,13</b>	<b>17,11</b>
<b>STDEV [MS/m]</b>	<b>0,00544</b>	<b>0,01144</b>	<b>0,00381</b>	<b>0,00406</b>	<b>0,01163</b>